

# BIOCOMPATIBILITY CERTIFICATE

**Testmaterial:** MoguCera C

**Supplier:** S&S Scheftner GmbH

Gonsenheimer Str. 56a, D-55126 Mainz

**Studies performed:** The following studies were performed in order to determine the biocompatibility of the product group implying MoguCera C according to ISO 10993-1:

**CYTOTOXICITY**

**SENSITISATION**

**INTRACUTANEOUS REACTIVITY**

**Results:** The product did not show any effect in the studies performed.

**BSL BIOSERVICE Scientific Laboratories GmbH Munich**

Behringstraße 6/8

D-82152 Planegg



Dr. Daniela Stelter

Biological Safety Testing

Date: November 09, 2011

